

FORM OF SHORT QUOTATION NOTICE

Government Engineering College Thrissur

D2/983/23/GECTCR

31.01.2023

Sealed quotations are invited for the supply of the Benchtop laboratory centrifuge for Chemical engg. department of this institution:

Seperate List attached

The envelopes containing the quotation should bear the superscription "Qtn. No.D2/86/22-23 due on **14/02/2023**" and should be addressed to the Principal, Govt. Engg. College, Thrissur, Kerala. Intending tenderers may submit the quotations on their own papers.

Last date for receipt of quotations is **14/02/2023 2.00pm** Late quotations will not be accepted. The quotations will be opened at **03.00 pm on 15/02/2023** in the presence of such of the tenderers or their authorized representatives who may be present at that time. The maximum period required for delivery of the articles should also be mentioned.

Details of the requirements and the conditions governing their supply can be obtained free on request from Principal, Govt. Engg. College, Thrissur, Kerala. till **14/02/2023 2.00pm**

Place:Thrissur

13.11.2019

Principal

copy to

1.Dr. Praseetha P Nair Che Engg Dept

2.Office

3.Notice board

| Sl. No. | Description/Specifications of the Item | No. of units required |
|---------|---|-----------------------|
| 1 | <p><u>Centrifuge</u></p> <p><u>Type</u></p> <ul style="list-style-type: none"> • Benchtop Laboratory centrifuge <p><u>Features</u></p> <ul style="list-style-type: none"> • Fixed angle head rotor • Stepless speed regulator • Digital speed indicator • Durability • Low maintenance • Dynamic brake for quick deceleration • 0–60-minute digital countdown timer • Safety lid interlock to prevent cover opening during centrifugation. • Dimension(mm) 330×370×295 • Shape rectangular • Non-corrosive stainless-steel material <p><u>Technical Specification</u></p> <ul style="list-style-type: none"> • Max. speed- 0 to 4400 rpm • Max. relative centrifugal force- 2350g • Max. capacity- 200ml <p><u>Rotor Head listing (Fixed Angular Type) with polypropylene tube</u></p> <ul style="list-style-type: none"> • Max. speed- 4400rpm • Max. relative centrifugal force- 2200g • Max. capacity- 8×15ml | 1 |

Approval Valid

Digitally Approved By
 Dr Ranjini Bhattathiripad T *
 Date: 31.01.2023
 Reason: Approved